

DECLARATION OF CONFORMITY

Certificate No 000007

GT

This declaration confirms that the above product, manufactured by

Gas Measurement Instruments Ltd
Inchinnan Business Park
Renfrew
Scotland
PR4 9RG

conforms to all the relevant Standards and Directives and is manufactured in accordance with Standards and Quality Assurance requirements.

The product is in compliance with the following Directives: -

Low Voltage Directive	73/23/EEC, 93/68/EC
ATEX Directive	ATEX 94/9/EC SIRA 05ATEX2043X
EMC Directive	93/68/EEC

The product has been tested to the following harmonised standards as appropriate to the Directives: -

EN50014:1997, EN50020:2002

The product complies with the following scheme: -

IEC IECEx SIR 05.006X

The product has been tested to the following standards: -

IEC 60079-0:2000, IEC 60079-1:2003, IEC 60079-11:1999, IEC 61779-1:1998, IEC 61779-4:1998, IEC 61779-5:1998

The quality is manufactured under an ISO9000:2000 quality system, BS EN 13980 and has product quality assurance surveillance as per the relevant Directives by

BSI ISO 9001:2008 - BSI Cert No Q0970
SIRA Module D - SIRA 00 ATEX M077 - SIRA Notified Body 0518

Gas Measurement Instruments Ltd confirms that the product and its associated manufacturing processes are in compliance with the above Directives and Standards.



Andrew Conway
Quality Manager

19th August'09